

This hospital intends to procure the following item as mentioned below:-

S. No.	Name of the item with specification	Quantity required
1.	<u>Heating and cooling dry bath with block</u> <u>Specifications:</u> -Instant Temperature display, both heating and cooling -Temperature range:- Below room temp i.e. 15 degree C to 100 degree C. -Temperature Stability: +0.5 degree C - Display Accuracy: 0.1 degree C - Heating Time: <12 min (25 degree C to 100 degree C) - Cooling Time: 20- 25 min (15 degree C to 0 degree C) - Should be multi- point operation - Should have LCD display and over temperature protection - Should have multipoint temperature operation, minimum 05 - Tube size: 35 X 2ml - Block capacity should be minimum 24 - Warranty: 5years - Yearly calibration with traceability certificate from NABL accredited laboratory. - Preventive maintenance services to be provided quarterly.	01 nos.

Those vendors who are ready to supply the item as per above-mentioned requirements are requested to submit their quotations in two parts (i.e. Technical bid (containing catalogue along with sample of the quoted equipment + GST regd. certificate) and Price bid (Basic Rate + applicable GST). Both the bids to be sealed in separate envelopes mentioning Technical bid and Price bid on the envelope then both sealed envelopes to be kept in a large sealed envelope super scribing "Quotation for Heating and cooling dry bath with block due on-22.05.2025".

The successful bidder will have to submit performance security in the form of FDR/Bank Guarantee in favour of Director, CNBC for 5% of the cost of the equipment in favour of Director, CNBC. The performance security shall remain valid for a period of sixty days beyond the date of completion of all contractual obligations including CMC period.

Your sealed quotation must reach the Procurement Branch, Room no. 4, Administrative block, CNBC, Geeta Colony, Delhi on or before 22.05.2025 upto 1:00 PM.

Note: In case the firm submitted the price in the technical bid or without a separate sealed envelope of technical & price bid, then their quotation will be summarily rejected.

Copy to: - Asstt. Programmer, CNBC with a request to upload the same on hospital website.

Procurement Officer
CNBC
PH: 011-21210213